

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS PO Box 1450 Alexandra, Virginia 22313-1450 www.wepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/634,498	08/05/2003	Randall Woods	32311-CIP2	7062	
33357 ADVANCED	7590 11/14/200 MEDICAL OPTICS, IN	EXAMINER			
1700 E. ST. A	NDREW PLACE		MATTHEWS, WILLIAM H		
SANTA ANA	, CA 92705		ART UNIT	PAPER NUMBER	
			3774		
			MAIL DATE	DELIVERY MODE	
			11/14/2008	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

# Office Action Summary

Application No.	Applicant(s)		
10/634,498	WOODS ET AL.		
Examiner	Art Unit		
William H. Matthews (Howie)	3774		

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS,

- WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.
- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed
- after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).

  Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any

earned	patent tern	n adjustment.	See 37	CFR	1./U4(b).	

Status
Responsive to communication(s) filed on 23 July 2008.  2a) ☐ This action is FINAL. 2b) ☐ This action is non-final.  3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.
Disposition of Claims
4) ⊠ Claim(s) 1-7.10 and 12-31 is/are pending in the application.  4a) Of the above claim(s) is/are withdrawn from consideration.  5) □ Claim(s) is/are allowed.  6) ☒ Claim(s) 1-7.10 and 12-31 is/are rejected.  7) □ Claim(s) is/are objected to.  8) □ Claim(s) are subject to restriction and/or election requirement.
Application Papers
9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are: a) control accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.
Priority under 35 U.S.C. § 119
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  a) All b) Some * c) None of:  1. Certified copies of the priority documents have been received.  2. Certified copies of the priority documents have been received in Application No.  3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).  * See the attached detailed Office action for a list of the certified copies not received.
Attachment(s)
1) Notice of References Cited (PTO-892)  4) Interview Summary (PTO-413)
2) Notice of Draftsperson's Patient Drawing Review (PTO-948) Paper No(s)/Mail Date.  3) Hoffmetton's Disclosure's Ottehement(s) (PTC/OB/K8) Ship Holice of Informal Patier L.Application  Paper No(s)/Mail Date.  1) Holice of Draftsperson's Patient Drawing Review (PTO-948)  5) Holice of Draftsperson's Patient Drawing Review (PTO-948)  1) Holice of Draftsperson's Patient Drawing Review (PTO-948)  2) Holice of Draftsperson's Patient Drawing Review (PTO-948)  2) Holice of Draftsperson's Patient Drawing Review (PTO-948)  2) Holice of Draftsperson's Patient Drawing Review (PTO-948)  3) Holice of Draftsperson's Patient Drawing Review (PTO-948)  4) Holice of Draftsperson's Patient Drawing Review (PTO-948)  4) Holice of Draftsperson's Patient Drawing Review (PTO-948)  5) Holice of Draftsperson's Patient Drawing Review (PTO-948)  6) Draftsperson's Patient Drawing Review (PTO-948)  1) Holice of Draftsperson's Patient Drawing Review (PTO-948)  1) Holice of Draftsperson's Patient Drawing Review (PTO-948)  1) Holice of Draftsperson's Patient Drawing Review (PTO-948)  2) Holice of Draftsperson's Patient Drawing Review (PTO-948)  2) Holice of Drawing Review (PTO-948)  3) Holice of Drawing Review (PTO-948)  4) Holice of Drawing Review (PTO-948)  5) Holice of Drawing Review (PTO-948)  5) Holice of Drawing Review (PTO-948)  6) Holice of Drawing Review (PTO-948)  6) Holice of Drawing Review (PTO-948)  6) Holice of Drawing Review (PTO-948)
S. Patent and Trademark Office  TOL-326 (Rev. 08-06) Office Action Summary Part of Paper No./Mail Date 20081110

Art Unit: 3774

#### DETAILED ACTION

#### Response to Arguments

Applicant's arguments with respect to claims 1-7,10,12-31 have been considered but are moot in view of the new ground(s) of rejection.

## Claim Objections

Claims 20 and 21 are objected to because of the following informalities: line 4 of claim 20 "an positioning member" and line 1 of claim 21 "the lens comprises plane passing through". Appropriate correction is required.

### Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 20-23,26,30-31 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention.

Claim 20, as amended, now recites less than all of the arms joined with legs at a bight. The specification do not disclose less than all being joined.

Application/Control Number: 10/634,498

Art Unit: 3774

Claims 27 and 29, depending from claim 1 which defines the opening between the segments (i.e. posterior and anterior segments), further describe structure extending across the optical axis. The specification does not describe an embodiment having posterior and anterior openings as well as structure meeting the limitations of claims 27 and 29.

Claims 21-23,26,30, and 31 depend from claim 20 and are rejected for including the base limitations described above.

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1-7,10,12-19 rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claims 1 and 16 define anterior and posterior segments, and further that the segments are "forming an opening disposed about the optical axis". It is unclear how "an opening" is formed by both the posterior and anterior segments when the optic is positioned therebetween. Examiner suggests amending claims 1 and 16 to conform with the appropriate language recited in lines 3-4 of claims 24-26 (i.e. a posterior opening and an anterior opening).

Claim 10 depends from cancelled claim 9.

Claims 2-7, 12-15, 17-19 depend from claims 1 or 16 and are rejected for including the base limitations described above.

Application/Control Number: 10/634,498

Art Unit: 3774

## Claim Rejections - 35 USC § 102

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filled in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filled in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-7,10,12-31 are rejected under 35 U.S.C. 102(e) as being anticipated by Portney USPN 7097660.

Portney disclose in figure 18 an intraocular lens comprising a resilient optic 350, optic positioning member 370,396,354,392,352,332,334 which is controlled by ciliary muscle action to distort the optic and comprises posterior/anterior segments surrounding, or "enclosing" the optic and providing openings as best understood. The positioning member comprises a pair of arms 392,396 and posterior/anterior segments are located on 352,354. Element 370 fulfills the limitations of claims 27,29,30. Portney disclose dimensions and powers (claims 13-15.17) in connection with figure 18.

#### Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to William H. Matthews (Howie) whose telephone number is 571-272-4753. The examiner can normally be reached on Monday-Friday 10-6:30PM.

Art Unit: 3774

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David J. Isabella can be reached on 571-272-4749. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/William H. Matthews/ Primary Examiner Art Unit 3774